

Ultrasonic cleaner HTM 01-05 Daily and Weekly tests Record sheet

Unit Serial No.	Customer reference and Location	Week commencing date
		/ /

Day	Cycle no.	PASS / FAIL	User no. and initials	End of day/session. Tank drained and decontaminated. Enter time	User no. and initials
Monday				:	
Tuesday				:	
Wednesday				:	
Thursday				:	
Friday				:	
Saturday				:	
Sunday				:	
				Load check test Insert day test is completed in 1 st column (under day heading)	
				Protein residue test Insert day test is completed in 1 st column (under day heading)	

Daily and weekly checks in brief

Daily: Visual inspection of instruments and record above.
Drain and decontaminate tank at end of day and record above.

Weekly: Load check – Record above and staple strip below.
Protein residue test – Record above and discard test.

Other tests
<p>In addition to the tests above, Quarterly and Annual tests should be performed. For further information please read leaflet “ultrasonic cleaner periodic testing advice”.</p>

<h2 style="margin: 0;">Staple load check strip here</h2>
--

If a printer is available, the printout for the above tests should be attached to this sheet. If the cycle information is captured electronically is should be recorded on the rear of this sheet. For further information please read leaflet "ultrasonic cleaner periodic testing advice" available online at www.walkerelectronics.com/downloads_advicesheets.php

Day	Cycle no.	User no.	Start time (hh:mm:ss)	Duration (seconds)	Start temp °C	End temp °C	Mean power (watts)	Power setting (0-100%)	Errors (detail or NONE)	Cycle PASS or FAIL
Mon			: : :							
Tue			: : :							
Wed			: : :							
Thur			: : :							
Fri			: : :							
Sat			: : :							
Sun			: : :							